

CITOFIX

IVD In vitro diagnostic medical device



Solution for fixation of cytological samples INSTRUCTIONS FOR USE

REF Product code: CX-1L (1000 mL)

Introduction

Impeccable sample fixation is a prerequisite for a correct cytological diagnosis. Tissue samples must be immersed in an optimally chosen fixative immediately after sampling, while they are still moist. Imminent fixation of the sample prevents drying and cell shape change, in turn enabling clear staining and a correct diagnosis. CitoFix solution is an ideal choice in liquid based cytology for treating hematology or cytology samples.

Product description

- **CITOFIX** - Solution for fixation of cytological samples. Contains an optimal concentration of alcohol, anticoagulants and preservatives.

Sampling and fixation

Using liquid based cytology method (LBC) and brush for collecting cytological samples, fixate the sample immediately by removing the brush head and immersing it in the fixative. At the beginning of processing the sample, isolate the cells from the fixative (one of the methods is to centrifuge the fixative) and place them on the microscope slide equally in a single layer. Cytological sample prepared in such a way is ready for staining.

Preparing the sample and diagnostics

Use only appropriate instruments for collecting and preparing the samples. Process the samples with modern technology and mark them clearly. Follow the manufacturer's instructions for use. In order to avoid mistakes, the staining procedure and diagnostics should only be conducted by authorized and qualified personnel. Use only microscope according to standards of the medical diagnostic laboratory. In order to avoid an erroneous result, a positive and negative check is advised before application.

Safety at work and environmental protection

Handle the product in accordance with safety at work and environmental protection guidelines. Used solutions and out of date solutions should be disposed of as special waste in accordance with national guidelines. Chemicals used in this procedure could pose danger to human health. Tested tissue specimens are potentially infectious. Necessary safety measures for protecting human health should be taken in accordance with good laboratory practice. Act in accordance with signs and warnings notices printed on the product's label, as well as in BioGnost's material safety data sheet which is available on demand.

Storing, stability and expiry date

Keep CitoFix in a tightly sealed original packaging at temperature between +15°C to +25°C. Do not freeze and avoid exposing to direct sunlight. Date of manufacture and expiry date are printed on the product's label.

References

1. Danos, M.L. (1974): *Fixatives for Cytologic Use*, 3rd ed., Chicago.
2. Keebler, C., Reagen J. (1983): *A Manual for Cytotechnology*, American Society of Clinical Pathologists Press, Chicago.
3. Conn, J. (1977): *Biological Stains*, 9th ed., Baltimore: Williams and Wilkens Co.

CX-1L, V4-EN3, 23.06.2020., IŠP/VR

	Refer to the supplied documentation		Storage temperature range		Number of tests in package	REF	Product code		European Conformity
	Refer to supplied instructions		Keep away from heat and sunlight		Valid until	LOT	Lot number		Manufacturer
IVD	For <i>in vitro</i> diagnostic use only		Keep in dry place		Caution - fragile				

BIOGNOST Ltd.
Medjugorska 59
10040 Zagreb
CROATIA
www.biognost.com

